



CERTIFICATE

of Conformity

EC Council Directive 2006/95/EC
Low Voltage Directive

Registration No.: ATS2014590

Applicant: PROKIT'S INDUSTRIES CO., LTD
5F, No 7, Lane 130, Ming-Chung Rd., Hsin-Tien Dist., New Taipei City,
Taiwan, R.O.C.

Product: 3 1/2 Digital LCR Multimeter

Identification:

| | | |
|------------|---|-------------------------------------------------------------------------------------------------------------------------------------------|
| Model No. | : | MT-5211 |
| Serial No. | : | n.a. |
| | | DC9V (powered by one 6F22 battery DC9V) Max. measuring voltage: 1000V DC/ 750V AC; Max. measuring AC/DC current: 20A (Max. 10 |
| Ratings | : | seconds of each 15 minutes); Measurement category: CAT IV 600V; Other function: Temperature, Resistance, Diode & Continuity etc. |

Standards: EN 61010-1:2010

The certificate of conformity is based on an evaluation of a sample of the above-mentioned product. Technical report and documentation are at the applicant's disposal. This is to certify that the tested sample is in conformity with all provisions of Annex I of Council Directive 2006/95/EC, referred to as the Low Voltage Directive. This certificate does not imply assessment of the production and does not permit the use of ATC's logo. The applicant of the certificate is authorized to use this certificate in connection with the EC declaration of conformity according to Annex III of the Directive.

Certified by

November 10, 2014

Date



Victor Chen

Victor Chen



The CE Marking may only be used if all relevant and effective EC Directives are complied with.



Accurate Technology Co., Ltd. - F1, Bldg. A&D, Changyuan New Material Port, Keyuan Rd., Science & Industry Park
Nanshan District, Shenzhen 518057, P.R. China
Tel: +86-755-26503290 Fax: +86-755-26503396 E-mail: webmaster@atc-lab.com Http://www.atc-lab.com